

X.XX x YY

NC ADVA GLIDE
PTCA Balloon Catheter - Non Compliant

REF

AN-DDD-LL

LOT

ANYYNNNN



MFG DATE

MM-YYYY




EXPIRY DATE

MM-YYYY

∅ Balloon: X.XX mm

↔ Balloon: YY mm



MIN. GUIDE CATHETER: 5F (0.058")

MAX. GUIDE WIRE DIAMETER: 0.014"

HYPOTUBE : 1.9 F



SHAFT PROX : 2.45 F

SHAFT DIST : 2.45 F

MONORAIL : 23 CM

TOT LENGTH : 140 CM

Balloon Compliance			
Pressure		Diameter	
Bar	kPa	mm	
4	400	x.xx	
6	600	x.xx	NP
8	800	x.xx	
10	1000	x.xx	
12	1200	x.xx	
14	1400	x.xx	
16	1600	x.xx	
18	1800	x.xx	
20	2000	x.xx	RBP
22	2200	x.xx	



STERILE EO

CE
2460

Label Ref: AN/LB/001/03

(01)XXXXXXXXXXXXXXXX(17)XXXXXX(10)ANXXXXXX
 NC ADVA GLIDE Advanced MedTech Solutions Private Limited AN-DDD-LL

(01)XXXXXXXXXXXXXXXX(17)XXXXXX(10)ANXXXXXX
 NC ADVA GLIDE Advanced MedTech Solutions Private Limited AN-DDD-LL

EC REP OBELIS S.A

Bd. General Wahis, 53
1030 Brussels, Belgium.
T: +32 2 732 59 54
F: +32 2 732 60 03
E: mail@obelis.net
W: www.obelis.net

M.R.P. Rs. XXXX.XX (Inclusive of all taxes)

X.XX x YY

NC ADVA GLIDE
PTCA Balloon Catheter - Non Compliant

REF AN-DDD-LL

LOT ANYYNNNN



EXPIRY DATE : MM-YYYY

Toll Free Number: 18004190607 email: customercare@amstld.com Mfg. License No: MFG/MD/2019/000174

*Technical Collaboration with
Conic Vascular Technology,
Switzerland*

Advanced MedTech Solutions Private Limited
P-21-22,25-26 & 34-35,GIDC Manjisar,
Tal-Savli, Dist. Vadodara-391775.
Manufactured in India by

