



<b>Declaration of Conformity</b>	
<b>Manufacturer:</b>	Volcano Corporation 2870 Kilgore Road Rancho Cordova, CA 95670 USA
<b>European Authorized Representative:</b>	Volcano Europe SA/ NV Excelsiorlaan 41 B-1930 Zaventem Belgium  Phone: +32 2 679 1076 Fax: +32 2 679 1079
<b>Conformity assessment procedure:</b>	93/42/EEC Annex II BSI Product Services Maylands Ave. Hemel Hempstead, Hertfordshire HP2 4SQ UK Notified Body Number 0086 Certificate No.: CE 86108
<b>Product:</b>	Revolution 45 MHz Intravascular Ultrasound Imaging Catheter Class III according to Annex IX rule 6 dash 1  GMDN Code: 44141  Product/Model #: 89000  CE Marking: 21 April 2006
<b>Harmonized Standards Applied:</b>	ISO 13485:2003, ISO 10993, ISO 11135, ISO 14971, ISO 11607
<p>I, the undersigned hereby declare that the medical device(s) specified above conform(s) to the essential requirements listed in Annex I of the European Medical Devices Directive, 93/42/EEC.</p>	
<b>Signature:</b>	
	24th November 2008
Ruud Nonnekens Manager RA/QA	Date