



4 Etikette

**Xpert**  
Stent System

REF LV\_ITNR LOT \*VOID

LV\_EXP

4 F  
1.33 mm

0.018"  
0.46 mm

120 cm

3 - 4 mm

20 mm

5.0 mm

**Xpert**  
Stent System

CONTENTS 1 Xpert OTW Delivery System with Self-Expanding Nitinol Stent

Ø STENT

**5.0 mm**

STENT

**20 mm**

SHEATH

**4 F**  
1.33 mm

REC GW

**0.018"**  
0.46 mm

Ø VESSEL

**3 - 4 mm**

USAB CAT

**120 cm**

REF LV\_ITNR LOT \*VOID LV\_EXP

CE 0123 ! 2 2 STERILE R

NON PYROGENIC

Abbott Laboratories Vascular Enterprises  
Anthoptstrasse  
8222 Beringen - Switzerland

Phone: +1 800 227-9902  
FAX: +1 800 601-8874  
Outside USA Phone: +1 951 914-4669  
Outside USA FAX: +1 951 914-2531

© 2011 Abbott Laboratories LA 008367001 005

| Rev. | Int. | PE       | Opt.               | Freig. | Beschreibung der Änderung                                                                                                      |
|------|------|----------|--------------------|--------|--------------------------------------------------------------------------------------------------------------------------------|
| 003  | MR   | NT       | BK, DK             | CLD    | Umstellung auf EAN Barcode                                                                                                     |
| 004  | MR   | R.Gerard | BK, ND             | CLD    | Rebranding + Umstellung auf Nice Label                                                                                         |
| 005  | STA  | BKU      | R. Jorgensen<br>IH | ESI    | Adressänderung Maastricht; Symbol „DIST“ entfällt und wird durch Customer Service Telefon & Faxnummern Santa Clara USA ersetzt |